

कार्यालय : असैनिक शल्य चिकित्सक सह मुख्य चिकित्सा पदाधिकारी, राँची।

सदर अस्पताल परिसर, राँची, email id – csofficeranchi@yahoo.com

Tender No. : 3957 Date 14.10.2023

पुर्न अति अल्पकालीन निविदा

निविदा संख्या 3683 दिनांक 3.10.2023 द्वारा जिला अस्पताल, राँची में Physiotherapy unit के मशीन उपकरण आदि के कय हेतु निबंधित निर्माणकर्ता/प्रतिष्ठानों/फर्म से दिनांक 13.10.2023 के अपराहन 5.00 बजे तक निविदा आमंत्रित किया जाता था परन्तु दो निविदा प्राप्त होने के कारण पुनः मशीन उपकरण एवं सामग्रियों के लिए अति अल्पकालीन निविदा दिनांक 25.10.2023 के अपराहन 5.00 बजे तक हाथो-हाथ अधोहस्ताक्षरी कार्यालय में प्राप्त किया जाएगा। निविदा संख्या 3683 दिनांक 3.10.2023 द्वारा जो निविदाताओं द्वारा निविदा समर्पित किया गया है उन्हें पुनः निविदा समर्पित करने की आवश्यकता नहीं है। निविदा से संबंधित तकनीकी विवरण/स्पेशिफिकेशन/संबंधित नियम/शर्त आदि website www.sadarhospital.com पर दिनांक 17.10.2023 से देखा जा सकता है। निविदा से संबंधित किसी प्रकार का परिवर्तन/शुद्धि पत्र www.sadarhospital.com पर ही देखा जा सकता है।

PR 308200 Health Edu and Family Welfare (23-24)D

असैनिक शल्य चिकित्सक सह
मुख्य चिकित्सा पदाधिकारी, राँची।

14.10.23

14.10.23

14/10/23

असैनिक शल्य चिकित्सक सह
मुख्य चिकित्सा पदाधिकारी, राँची।

Physiotherapy unit के निविदा का तकनीकी भाग :-




निविदा के तकनीकी भाग में निम्नांकित अभिलेख समर्पित करना अनिवार्य होगा।

1. जी०एस०टी० रजिस्ट्रेशन प्रमाण पत्र। (G.S.T Registration Certificate)
2. जी०एस०टी० का रिटर्न प्रतिवेदन माह अगस्त 2023 का संलग्न करना होगा।
3. आयकर कार्ड (पैन कार्ड) (Pan Card by Income Tax Department)
4. निर्माता कंपनी/ प्राधिकृत विक्रेता, निविदादाता एवं एजेंसी को इस आशय का शपथ पत्र/Undertaking देना होगा कि (a)उनकी कंपनी/संस्थान केन्द्र सरकार, राज्य सरकार या किसी अनुषांगीक इकाई द्वारा काली सूची में नहीं है (b) MRP से अधिक दर अंकित नहीं है
5. निविदादाता के द्वारा वित्तीय वर्ष, 2019-20, 2020-21 & 2021-2022 का आयकर रिटर्न प्रमाण पत्र संलग्न करना होगा।
6. निविदा के साथ वित्तीय वर्ष 2019-20, 2020-21 & 2021-2022 के प्रत्येक वर्ष 50,00,000.00 (पचास लाख) का टर्न ओवर होना चाहिए। इससे संबंधित सी.ए. का प्रमाण पत्र संलग्न करना होगा।
7. Produce Technical Specification along with Company Name with Make – Example Given Below :-

क्र०सं०	सामानों का नाम	निर्माता कंपनी का नाम	निर्माता कंपनी का ISO Certificates	अभियुक्ति
1				

अन्य शर्त :-

1. बिना कारण बतायें निविदा को आंशिक या पूर्ण रद्द करने का सर्वाधिकार जिला क्रय समिति को सुरक्षित रहेगा।
2. इस निविदा के आधार पर निर्धारित दर कम से कम एक वर्ष के लिए मान्य होगा।
3. मूल्य सूची में स्पेसिफिकेशन, मेक, ईकाई स्पष्ट रूप से अंकित होना चाहिए।
4. किसी प्रकार का विवाद होने की स्थिति में न्यायिक क्षेत्र रॉची होगा।
5. निविदा निविदा के द्वारा ही मान्य होगा।
6. सामानों आदि की सूची के क्रम संख्या के अनुसार ही सभी करें सहित मूल्य अंकित करेंगे।
7. क्रयदेश प्राप्ति के 15 दिनों अंदर सामानों आदि की आपूर्ति करना होगा तथा निर्धारित समय के अन्तराल में आपूर्ति नहीं होने की स्थिति में एक सप्ताह के बाद जिला क्रय समिति द्वारा काली सूची में डालने, जमानत राशि जब्त करने एवं निविदा में भाग लेने पर प्रतिबंध की कार्रवाई की जाएगी या जिला क्रय समिति द्वारा निर्णय के आलोक में जो दण्ड पारित होगा।
8. जिला क्रय समिति का निर्णय अन्तिम निर्णय होगा।
9. इस निविदा से संबंधित कोई भी शुद्धी पत्र या आदेश इसी वेब साइट www.sadarhospital.com पर अपलोड किया जाएगा।
10. दिये गये स्पेसिफिकेशन के अनुसार सामानों की आपूर्ति नहीं किये जाने पर किसी भी विवाद के लिए जिम्मेवारी संबंधित आपूर्तिकर्ता की होगी तथा भुगतान की कार्रवाई रोक दी जाएगी।


14.10.23

14.10.23

14.10.23

Annexure (मशीन उपकरण)

Sl. No	Uploaded document	Page number						
1	जी0एस0टी0 रजिस्ट्रेशन प्रमाण पत्र। (G.S.T Registration Certificate)							
2	जी0एस0टी0 का रिटर्न प्रतिवेदन माह अगस्त 2023 का संलग्न करना होगा।							
3	आयकर कार्ड (पैन कार्ड) (Pan Card by Income Tax Department)							
4	निर्माता कंपनी/ प्राधिकृत धिकेता, निविदादाता एवं एजेंसी को इस आशय का शपथ पत्र/Undertaking देना होगा कि (a)उनकी कंपनी/ संस्थान केन्द्र सरकार, राज्य सरकार या किसी अनुषांगीक इकाई द्वारा काली सूची में नहीं है (b) MRP से अधिक दर अंकितनहीं है							
5	निविदादाता के द्वारा वित्तिथि वर्ष 2019-20, 2020-21 & 2021 - 2022 का आयकर रिटर्न प्रमाण पत्र संलग्न करना होगा।							
6	निविदा के साथ वित्तीय वर्ष 2019-20, 2020-21 & 2021-2022 के प्रत्येक वर्ष 50,00,000.00 (पचास लाख) का टर्न ओवर होना चाहिए । इससे संबंधित सी.ए. का प्रमाण पत्र संलग्न करना होगा।							
7	निविदा से संबंधित निविदा E.M.D. 10000.00 (Ten Thousand) की राशि का मूल Bank D.D. निर्धारित तिथि तक कार्यालय में जमा कर देना होगा। तत्पश्चात निविदा मान्य होगा।							
8	Produce Technical Specification along with Company Name with Make – Example Given Below :- <table><tr><td>क्र०सं०</td><td>मशीन का नाम</td><td>निर्माता कंपनी का नाम</td></tr><tr><td>1</td><td></td><td>A/B/C/D</td></tr></table>	क्र०सं०	मशीन का नाम	निर्माता कंपनी का नाम	1		A/B/C/D	
क्र०सं०	मशीन का नाम	निर्माता कंपनी का नाम						
1		A/B/C/D						
9	निर्माता कंपनी का नाम	Machine Upkaran Name	निर्माता कम्पनी का ISO	Page No.	Remarks			

मशीन उपकरण की सूची , स्पेशिफिकेशन एवं बी.ओ.क्यू

क सं	मशीन उपकरण का नाम	Make	Unit	Price Including All Taxes & Others.
1	Cryo therapy - Technical Specifications CRYO Therapy (CRYO) Cryo therapy issue of extremely low temperatures mainly used in sports medicine for faster pain relief and edema reduction. It has been scientifically proven to bring therapeutic effects in rehabilitation, mainly to eliminate inflammation, pain and edema, and reduce spasticity. Most common indications being pain reduction, inflammation, muscle spasm, sports injuries, post-surgery conditions & rheumatic diseases <ul style="list-style-type: none"> The unit should have compact design & silent operation System should come with maintenance free features (no water tank) The device must provide air flow of 350-1200 liters / minute Operating Air temperature must be up to -32 degrees Celsius The Therapy Hose should be light weight and min 1 mtrs in length The Unit should be supplied with the minimum 3 sizes of nozzles for different anatomic sites i.e. 5 mm, 10 mm & 15 mm diameter Option for Wide angled nozzle should be available The device should have option for therapy arm support holder Accessories of glass plate for holding hose and applicators The device must be supplied with built in Trolley with castors Mains supply 100-240 V, 50-60 Hz Should be European CE certified / International Quality controls One year warranty & after warranty two years AMC 		Each	

14.10.23

14.10.23

14.10.23

2	<p>Shockwave Therapy - General Characteristics:</p> <ol style="list-style-type: none"> 1. Operating Pressure –upto 4 Bars 2. System must have upgrade option to go to 5 bars 3. Frequency – upto 15Hz 4. System must have option to increase frequency up to 20 Hz 5. Highly-effective, non-invasive treatment for pain associated with musculoskeletal system 6. Built-in Therapeutic Encyclopaedia – description and images explaining application of all pre-programmed treatment programs 7. Should have pre-programmed protocols 8. Large touch screen display - 5.7 inches 9. Self-identifications of accessories and accessories 10. User defined programmes also should be available 11. Setting of sound effects, display, auto turn-off 12. Intensity Gradient Model should be available 13. Applicator must eliminate backward shocks 14. Applicator must give at least 20,00,000 shocks 15. User defined programmes also should be available 16. Mains supply 100-240V / 50-60 Hz 17. Class according to MDD 93/42/EEC IIB 18. Must be supplied with Trolley with castors & minimum 4 selves for storing accessories <p>Should be supplied with Standard Accessories</p> <p>Main unit with in-built air compressor, SWT applicator, 300ml, SWT applicator holder, touch screen pen point</p> <p>One year warranty & after warranty two years AMC</p>		Each	
3	<p><u>LaserTherapy</u> Should have medical encyclopaedia into the system with anatomical library, and informative screens There should be facility of changing treatment parameters and storing them into user/patient file Should have large colour touch screen display; at least 7 inch Should have large number of pre-programmed indications (more than 50) Should have facility for pre- programmed protocol selection through list navigation/quickprotocols/ body part navigation. Should have the facility of up-grading the firmware in future. Should have facility to connect 2 laser applicators at the same time and same should be available for selection while setting up parameters for treatment through software saving operator's time and effort. Should have option of cluster laser probe as well as point probe.</p>		Each	

14.10.23
14.10.23

14.10.23

	<p>Should have option for operation through battery. Continuous and pulse therapy options Automatic recalculation of therapy parameters Laser acupuncture and Nogier's frequencies Laser probe should have navigational beam Option for battery operation Device must be supplied with trolley having castors and at least 6 drawers Technical features:</p> <ul style="list-style-type: none"> • Number of probes simultaneously connected: 2 • Modes of operation: Continuous and pulse • Dosage Fine: 0.1–99.0 J / cm² • Frequency: 0–10000 Hz for mono probe : 0–500 for cluster probe • Therapy area: 0.1–50 cm² • Duty factor: 10–90 % • Laser class 3B • Dimension 380 × 190 × 260 mm • Average power of mono probe in 830 nm wavelength with options of Power of 50 mW / 100 mW / 200mW / 300 mW / 400 mW • Average power of cluster laser probe in options of 200 mW / 800mW / 1000 mw / 1300 mW / 1500 mW <p>One year warranty & after warranty two years AMC</p>			
4	<p>Compression Therapy. The device should work on Pneumatic Presso therapy principle which provides gentle massage that encourages through the body. It has been scientifically as well as clinically proven to bring therapeutic effects such as muscle relaxation, edema reduction, increase circulation and supporting overall faster rehabilitation from injury / support tissue healing General features: Number of Chambers should be 12 Working pressure should be 20 - 160mmHg The pressure should be adjustable for each chamber/segment separately Gradient: 0-100%, in steps of 1% It must have touch screen display; min 5.7 inch screen Should have at least 15 pre-programmed protocols Should have at least 100 memory spaces for user defined Must have built in Encyclopaedia with treatment recommen Should be provided with Trousers with 24 chambers with zipper for easy and quick dressing It should have Leg cuff with 10chambers and Arm cuff with 8 chambers available as options in medium</p>	Each		

14.10.23
14/10/23

14.10.23

Optional Interface for using 2 arms or two legs together			
Working chamber activity should be shown on display			
There should be option to disable one or more chambers. if patient cannot tolerate pressure			
Trolley with Castors must be provided with at least 4-5 drawers			
One year warranty & after warranty two years AMC			

Annexure (सामग्री)

Annexure (समीक्षा)					Page number
Sl. No	Uploaded document				
1	जी0एस0टी0 रजिस्ट्रेशन प्रमाण पत्र। (G.S.T Registration Certificate)				
2	जी0एस0टी0 का रिटर्न प्रतिवेदन माह अगस्त 2023 का संलग्न करना होगा।				
3	आयकर कार्ड (पैन कार्ड) (Pan Card by Income Tax Department)				
4	निर्माता कंपनी/ प्राधिकृत विक्रेता, निविदादाता एवं एजेंसी को इस आश्वासन पत्र/Undertaking देना होगा कि (a)उनकी कंपनी/संस्थान केन्द्र सरकार, राज्य सरकार या किसी अनुषांगीक इकाई द्वारा काली सूची में नहीं है (b) MRP से अधिक दर अंकित नहीं है				
5	निविदादाता के द्वारा वित्तिय वर्ष 2020-21 , 2021 – 2022 , 22-23 का आयकर रिटर्न प्रमाण पत्र संलग्न करना होगा।				
6	निविदा के साथ वित्तीय वर्ष, 2020-21 , 2021-2022,22-23 के एमरेज वार्षिक टर्न ओवर 5,00,000.00 (पाँच लाख) का होना चाहिए । इससे संबंधित सी.ए. का प्रमाण पत्र संलग्न करना होगा।				
7	निविदा से संबंधित निविदा E.M.D. 10000.00 (Ten Thousand) की राशि का मूल Bank D.D. निर्धारित तिथि तक कार्यालय में जमा कर देना होगा। तत्पश्चात निविदा मान्य होगा।				
8	Produce Technical Specification along with Company Name with Make – Example Given Below :-				
	क्र०सं०		मशीन का नाम	निर्माता कंपनी का नाम	
	1			A/B/C/D	
9	निर्माता कंपनी का नाम	Item Name	Make	Page No.	Remarks

आयकर की सूची स्पेशिफिकेशन एवं बी.ओ.क्यू

सामग्रियों की सूची, स्पेशिफिकेशन एवं बी.ओ.क्यू

क्र सं	सामग्रियों का नाम	Make if any	Unit	Price Including All Taxes & Others.
1	Folder Printing - 300GSM Art Board with multi colour.			
2	Nursing Station With Table - per sq fit (Specification - Aluminum Panel, Aluminum Corcated Sheet , 6 mm Glass , 18 mm water proof ply with sunmica work .			
3	Shed Work Per Sq fit - Specification - Crop Sheet , 2/2 inch holo sq pipe , 1.5/1.5 inch holo sq pipe , Sheetting bolt , Earth Excavations, 1.5-1 fit concrete work			
4	Monitor Stand - per sq inch - MDF ply board Fix with iron stand .			
5	False Ceiling Work - per Sq fit (Specification - Gypsum board with iron panel)			
6	Roller Parda Specification Attached			
7	Water Fountain work with Color full light - Specification Attached			
8	Patient Trolley with Steel railing, Hydraulic, Oxy Cylinder Stand , Sline Stand, Mattress , best Quality	Each		
9	Three Rack Steel food Trolley	Each		
10	Water proofing Work per Sq fit			
11	Aluminum partition work with ACP - per Sq fit			
12	Bio metric machine - Wall Hanging	Each		
13	Battery - AAA	Each		
14	Button Coin cell For Glucometer/Heamoglobino meter.	Each		

14-10-23
14-10-23

14-10-23

Sr	Description of Item	unit	Qty
1.	STRAIGHT CEILING RAIL. PROFIL 25*25 mm. ALUMINIUM ANODIZED LOAD CAPACITY 50 kg	NOS	1
2.	END PIECE LOCK FOR THE RAIL SYSTEM. MATERIAL ALUMINIUM	NOS	1
3.	CABLE CARRIER TRACK RUNNER	NOS	1
4.	HOSPITAL TRACK. MATERIAL ALUMINIUM ALLOY, WHITE COLOUR. LENGTH OF TRACK AS PER REQUIRED UP TO 16 FEET.	NOS	1
5.	COVER. MATERIAL POLYPROPYLENE & STEEL, WHITE COLOUR	NOS	1
6.	MASTER RUNNER. MATERIAL POLYPROPYLENE, WHITE COLOUR	NOS	1
7.	T-SHOPE CONNECTOR. MAGNETIC STEEL, 6 MAGNETIC SCREW, SILVER COLOUR	NOS	1
8.	CONNECTOR. ALUMINIUM ALLOY. (I) 2 MAGNETIC STEEL PLATE (II) 6 MAGNETIC STEEL SCREW	NOS	1
9.	HEANGING ROD (FOR HANGING). ALUMINIUM ALLOY, SILVER COLOUR, WEIGHT PER RUNNING METER 37.02G	NOS	1
10.	UPPER BAR (FOR HANGING) ALUMINIUM ALLOY, 2 HEXAGON SCREW, SILVER COLOUR	NOS	1
11.	BOTTOM BAR (FOR HANGING) ALUMINIUM ALLOY, CONTAINS (I) 2 HEXAGON SCREW (II) 1 MAGNETIC STEEL BOLT. SILVER COLOUR, WEIGHT 33.78/PC	NOS	1
12.	LOCKING PLATE. MAGNETIC STEEL, COLOUR SILVER	NOS	1

Water Fountain work with Color full light - SI no - 7

SL	Description of item
01.	10 FEET DIA Ring make in Jindal B CLASS grade Complete Holder assembly with Rubber coated paint.
02.	Clear Stream Nozzle With Ball Joint Make in Brass with shine chrome
03.	Vertical Center Jet Nozzle With Ball Joint Make in Brass with shine chrome
04.	Programmable Underwater Color Changing LED Light 18 W (powered By Edison)
05.	DCC Programmable Controller for Light Mode Model-DPS/PWN
06.	Open well submersible pump with dust grill 3HP, pumps are according to suitable design Make : Crompton / Kirloskar / CRI
07.	Electrical panel 16 Gage CRC Sheet, power coated, contractor overload relay with face and on/off indication, protection circuit. Assembled As Suitable for Pump Operational Systems as water feature Required
08.	Supplying and Fixing of PVC/UPVC pipes (10 Kg/cm2) of make Prince/Supreme/Prakash or equivalent confirming including all UPVC fitting / jointing with good quality solvent.
09.	Supplying and Laying submersible flat/round cables PVC insulated copper conductor of complete as per the direction of Company. Thickness of wire 1.5mm, 2.5mm. Jointing of wire PVC. Waterproof coating tape or chemical. Note- All wires are in 100% copper
10.	Drawing/Design/Testing /Supervision / Installation
11.	Granite & Tilling Work
12.	Waterproofing of water-storage part, Civil work with Plumbing work

असैनिक शल्य चिकित्सक सह
मुख्य चिकित्सा प्रदायिकारी, रॉबी।

14-10-23 11:10:23
14/10/23